

### Cleaning

- Unplug the VorOtek Scope or Light from power source prior to cleaning.
- Visually inspect the VorOtek Scope or Light for soiling.
- Remove all visible soiling by first wiping external surfaces & cables with a soft cloth slightly dampened with mild detergent and water (mixed to the manufacturer's instructions).
- Wipe all external surfaces, including Headband model leather covers & velcro straps using a soft cloth slightly dampened with 70% isopropyl Alcohol (IPA).
- Allow all surfaces to air dry for at least 60 seconds.
- Re-wipe all surfaces with a lint free soft cloth to remove any residue from cleaning products.
- Ensure all surfaces & cables are dry before operating the VorOtek Scope or Light.
- Caution: Do not immerse or saturate any component of the VorOtek Scope or Light in a cleaning solution.
- Caution: Do not autoclave any component of the VorOtek Scope or Light.

### Maintenance

- Before every procedure, carefully inspect the VorOtek Scope or Light to ensure it is fully functional. DO NOT use if inspection reveals any damage.
- The VorOtek Scope & Light Range does not require any other routine maintenance.
- For repair or warranty, refer to the support & repair page on VorOtek's website.

### Warranty

- **Scopes & Lights** - 3 years for any component defect or manufacturing failure.
- **Power Pack** - 18 months for any component defect or manufacturing failure.
- Warranty does not cover abuse, misuse or transit damage.
- Please contact VorOtek prior to sending for repair.

### Notices & Precautions

- Should a serious incident occur while using the device please report the incident to the manufacturer (VorOtek) and the competent authority of the Member State in which the user and /or patient is established.
- Only to be used by qualified health professionals.
- Do not use in the presence of flammable anaesthetics
- Do not expose electrical equipment to liquids or excessive moisture.



## VOROTEK

23 Point Henry Road  
MOOLAP VIC 3224, Australia  
T +61 3 5248 5809  
E [office@vorotek.com.au](mailto:office@vorotek.com.au)  
[www.VorOtek.com](http://www.VorOtek.com)



EMERGO EUROPE  
Prinsessegracht 20  
2514 AP The Hague  
The Netherlands



IFU-007-HeadLight-Fixed Spot TechMan-R5

# LED Headlight

FIXED SPOT

PORTABLE SURGICAL QUALITY LED ILLUMINATION

## Technical Manual

# Technical Manual

## LED Illumination

- High efficiency LED (Light Emitting Diodes) rated at 50,000 hours operating life.
- Light Intensity: 80,000 Lux.
- Colour Temperature: 5,700 Degrees Kelvin (Daylight).
- Colour Rendering Index: CRI 75.
- Caution: Hot surface. Avoid direct skin contact with aluminium heat sink during extended periods of use.
- Caution: Do not look directly into the LED Light.

## Dimmable Illumination

The dimmer switch is located on the cable near the Power Pack Plug. The dimmer function provides 8 levels of illumination with "H" being the highest. (Refer Figure 1)

**Dimming:** Press the "+" or "-" button on the dimmer switch to select the preferred level of illumination.

**ON/OFF function:** Click & hold either button on the dimmer switch for 2 seconds to turn illumination "on" or "off" at the selected level.

Figure 2. Dimmer Switch



## Power Pack

- Power Packs are Lithium Ion: 11.1 Volts, 2,900 mAh.
- The Power Pack delivers 7hrs to >20hrs "on time" depending on illumination level.
- The Power Pack features short circuit protection and charging control circuitry.
- Two position switch - ON/OFF.
- Designed to be worn in the breast pocket or to be clipped to a belt. It can also be clipped to the back of a surgical gown.
- Caution: Do not expose the Power Pack to heat, fire or mechanical shock.
- Caution: Do not open or dismantle the Power Pack.
- Caution: Only use approved Power Packs supplied by Vorotek.

## Charger - Charging Power Pack

- Connect Power Pack to Charger.
- The Power Pack can be charged in either the ON or OFF position.
- RED INDICATORS – Charging (takes up to 5 hours).
- GREEN INDICATORS – 100% Charged.
- Charging is best done overnight, frequently and before the Power Pack runs flat.
- Caution: Use the supplied 12.6V Charger only.
- Caution: Do not attempt to operate the Charger if the cord and/or plug appear to be damaged.
- Caution: Do not use Charger in the patient vicinity.

## HeadBand Bracket Adjustment

The HeadBand mounting bracket has 3 adjustable links allowing a full range of adjustment i.e. the angle of declination and forward position of the illumination system.

## Troubleshooting

### LED illumination cuts out and resets repeatedly

- The LED illumination does not fade, but the protective circuit breaker keeps switching off (and resetting) when the Power Pack requires recharging. See section on 'Charger'.
- LED's rarely need replacing, but if this is necessary it must be done by Vorotek.

### Power Pack will not charge

- See section on 'Charger'.

**For all other questions contact Vorotek.**